



ANNEX A

CALIBRATION CERTIFICATES



REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION

CONTACT: JOHNNY HO
CLIENT: EGS (ASIA) LTD
ADDRESS: 15/F., NORTH POINT INDUSTRIAL BUILDING,
499 KING'S ROAD, NORTH POINT,
HONG KONG

WORK ORDER: HK2329342
SUB-BATCH: 0
LABORATORY: HONG KONG
DATE RECEIVED: 25-Jul-2023
DATE OF ISSUE: 26-Jul-2023

SPECIFIC COMMENTS

Equipment information (Brand name, Model No., Serial No. and Equipment No.) is provided by client. The performance of the equipment stated in this report is checked with independent reference material and results compared against a calibrated secondary source.

The "Tolerance Limit" quoted is the acceptance criteria applicable for similar equipment used by the laboratory or quoted from relevant international standards.

The "Next Calibration Date" is recommended according to best practice principle as practised by the laboratory or quoted from relevant international standards.

The validity of equipment/ meter performance only applies to the result(s) stated in the report.

Equipment Type: Multifunctional Meter
Service Nature: Performance Check
Scope: Dissolved Oxygen, pH Value, Turbidity, Salinity and Temperature
Brand Name/ Model No.: [YSI]/ [6820-V2-M]
Serial No./ Equipment No.: [07H100241]/ [MPP22]
Date of Calibration: 25-July-2023

GENERAL COMMENTS

This report superseded any previous report(s) with same work order number.

Ms. Lin Wai Yu, Iris
Assistant Manager - Inorganics

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION



WORK ORDER: HK2329342
SUB-BATCH: 0
DATE OF ISSUE: 26-Jul-2023
CLIENT: EGS (ASIA) LTD

Equipment Type: Multifunctional Meter
Brand Name/ Model No.: [YSI]/ [6820-V2-M]
Serial No./ Equipment No.: [07H100241]/ [MPP22]
Date of Calibration: 25-July-2023

Date of Next Calibration: 25-October-2023

PARAMETERS:

Dissolved Oxygen

Method Ref: APHA (23rd edition), 4500O: G

Expected Reading (mg/L)	Displayed Reading (mg/L)	Tolerance (mg/L)
2.96	2.80	-0.16
5.49	5.43	-0.06
7.18	7.19	+0.01
	Tolerance Limit (mg/L)	±0.20

pH Value

Method Ref: APHA (23rd edition), 4500H: B

Expected Reading (pH unit)	Displayed Reading (pH unit)	Tolerance (pH unit)
4.0	3.91	-0.09
7.0	6.99	-0.01
10.0	9.89	-0.11
	Tolerance Limit (pH unit)	±0.20

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

Ms. Lin Wai Yu, Iris
Assistant Manager - Inorganics

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION



WORK ORDER: HK2329342
SUB-BATCH: 0
DATE OF ISSUE: 26-Jul-2023
CLIENT: EGS (ASIA) LTD

Equipment Type: Multifunctional Meter
Brand Name/ Model No.: [YSI]/ [6820-V2-M]
Serial No./ Equipment No.: [07H100241]/ [MPP22]
Date of Calibration: 25-July-2023

Date of Next Calibration: 25-October-2023

PARAMETERS:

Turbidity

Method Ref: APHA (23rd edition), 2130B

Expected Reading (NTU)	Displayed Reading (NTU)	Tolerance (%)
0	0.0	--
4	4.3	+7.5
40	40.7	+1.8
80	81.4	+1.8
400	N/A	N/A
800	N/A	N/A
Tolerance Limit (%)		±10.0

Salinity

Method Ref: APHA (23rd edition), 2520B

Expected Reading (ppt)	Displayed Reading (ppt)	Tolerance (%)
0	0.00	--
10	9.73	-2.7
20	19.46	-2.7
30	29.65	-1.2
Tolerance Limit (%)		±10.0

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

Ms. Lin Wai Yu, Iris
Assistant Manager - Inorganics

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION



WORK ORDER: HK2329342
SUB-BATCH: 0
DATE OF ISSUE: 26-Jul-2023
CLIENT: EGS (ASIA) LTD

Equipment Type: Multifunctional Meter
Brand Name/ Model No.: [YSI]/ [6820-V2-M]
Serial No./ Equipment No.: [07H100241]/ [MPP22]
Date of Calibration: 25-July-2023

Date of Next Calibration: 25-October-2023

PARAMETERS:

Temperature

Method Ref: Section 6 of International Accreditation New Zealand Technical Guide No. 3 Second edition March 2008: Working Thermometer Calibration Procedure.

Expected Reading (°C)	Displayed Reading (°C)	Tolerance (°C)
11.9	12.01	+0.1
20.7	19.75	-0.9
39.5	38.79	-0.7
	Tolerance Limit (°C)	±2.0

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

Ms. Lin Wai Yu, Iris
Assistant Manager - Inorganics



REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION

CONTACT: DOMINIC LAI
CLIENT: EGS (ASIA) LTD
ADDRESS: 15/F., NORTH POINT INDUSTRIAL BUILDING,
499 KING'S ROAD, NORTH POINT, HONG KONG

WORK ORDER: HK2340210
SUB-BATCH: 0
LABORATORY: HONG KONG
DATE RECEIVED: 10-Oct-2023
DATE OF ISSUE: 12-Oct-2023

GENERAL COMMENTS

The performance of the equipment stated in this report is checked with independent reference material and results compared against a calibrated secondary source.

The "Tolerance Limit" quoted is the acceptance criteria applicable for similar equipment used by the laboratory or quoted from relevant international standards.

The "Next Calibration Date" is recommended according to best practice principle as practised by the laboratory or quoted from relevant international standards.

The validity of equipment/ meter performance only applies to the result(s) stated in the report.

This report superseded any previous report(s) with same work order number.

EQUIPMENT INFORMATION

Equipment information (Brand name, Model No., Serial No. and Equipment No.) is provided by client.

Equipment Type: Multifunctional Meter

Service Nature: Performance Check

Scope: Dissolved Oxygen, pH Value, Turbidity, Salinity and Temperature

Brand Name/ Model No.: [YSI]/ [6920-V2-M]

Serial No./ Equipment No.: [08C100240]/ [MPP30]

Date of Calibration: 10-October-2023

Ms. Lin Wai Yu, Iris
Assistant Manager - Inorganics

This report may not be reproduced except with prior written approval from ALS Technichem (HK) Pty Ltd.

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION



WORK ORDER: HK2340210
SUB-BATCH: 0
DATE OF ISSUE: 12-Oct-2023
CLIENT: EGS (ASIA) LTD

Equipment Type: Multifunctional Meter
Brand Name/ Model No.: [YSI]/ [6920-V2-M]
Serial No./ Equipment No.: [08C100240]/ [MPP30]
Date of Calibration: 10-October-2023 Date of Next Calibration: 10-January-2024

PARAMETERS:

Dissolved Oxygen Method Ref: APHA (23rd edition), 4500O: G

Expected Reading (mg/L)	Displayed Reading (mg/L)	Tolerance (mg/L)
3.06	3.09	+0.03
4.04	4.19	+0.15
7.26	7.18	-0.08
	Tolerance Limit (mg/L)	±0.20

pH Value Method Ref: APHA (23rd edition), 4500H: B

Expected Reading (pH unit)	Displayed Reading (pH unit)	Tolerance (pH unit)
4.0	3.88	-0.12
7.0	7.01	+0.01
10.0	9.82	-0.18
	Tolerance Limit (pH unit)	±0.20

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

Ms. Lin Wai Yu, Iris
Assistant Manager - Inorganics

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION



WORK ORDER: HK2340210
SUB-BATCH: 0
DATE OF ISSUE: 12-Oct-2023
CLIENT: EGS (ASIA) LTD

Equipment Type: Multifunctional Meter
Brand Name/ Model No.: [YSI]/ [6920-V2-M]
Serial No./ Equipment No.: [08C100240]/ [MPP30]
Date of Calibration: 10-October-2023

Date of Next Calibration: 10-January-2024

PARAMETERS:

Turbidity

Method Ref: APHA (23rd edition), 2130B

Expected Reading (NTU)	Displayed Reading (NTU)	Tolerance (%)
0	0.0	--
4	3.7	-7.5
40	43.3	+8.2
80	84.3	+5.4
	Tolerance Limit (%)	±10.0

Salinity

Method Ref: APHA (23rd edition), 2520B

Expected Reading (ppt)	Displayed Reading (ppt)	Tolerance (%)
0	0.00	--
10	9.55	-4.5
20	19.38	-3.1
30	28.78	-4.1
	Tolerance Limit (%)	±10.0

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

Ms. Lin Wai Yu, Iris
Assistant Manager - Inorganics

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION



WORK ORDER: HK2340210
SUB-BATCH: 0
DATE OF ISSUE: 12-Oct-2023
CLIENT: EGS (ASIA) LTD

Equipment Type: Multifunctional Meter
Brand Name/ Model No.: [YSI]/ [6920-V2-M]
Serial No./ Equipment No.: [08C100240]/ [MPP30]
Date of Calibration: 10-October-2023 Date of Next Calibration: 10-January-2024

PARAMETERS:

Temperature

Method Ref: Section 6 of International Accreditation New Zealand Technical Guide No. 3 Second edition March 2008: Working Thermometer Calibration Procedure.

Expected Reading (°C)	Displayed Reading (°C)	Tolerance (°C)
14.5	14.20	-0.3
21.0	21.58	+0.6
40.5	40.27	-0.2
	Tolerance Limit (°C)	±2.0

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

Ms. Lin Wai Yu, Iris
Assistant Manager - Inorganics



ALS Technichem (HK) Pty Ltd

11/F., Chung Shun Knitting Centre,

1 - 3 Wing Yip Street,

Kwai Chung, N.T., Hong Kong

T: +852 2610 1044

F: +852 2610 2021

www.alsglobal.com

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION

CONTACT: LAM MEI SHING
CLIENT: EGS (ASIA) LTD
ADDRESS: 15/F., NORTH POINT INDUSTRIAL BUILDING,
499 KING'S ROAD, NORTH POINT, HONG KONG

WORK ORDER: HK2346638
SUB-BATCH: 0
LABORATORY: HONG KONG
DATE RECEIVED: 21-Nov-2023
DATE OF ISSUE: 22-Nov-2023

GENERAL COMMENTS

The performance of the equipment stated in this report is checked with independent reference material and results compared against a calibrated secondary source.

The "Tolerance Limit" quoted is the acceptance criteria applicable for similar equipment used by the laboratory or quoted from relevant international standards.

The validity of equipment/ meter performance only applies to the result(s) stated in the report.

This report superseded any previous report(s) with same work order number.

EQUIPMENT INFORMATION

Equipment information (Brand name, Model No., Serial No. and Equipment No.) is provided by client.

Equipment Type: Multifunctional Meter

Service Nature: Performance Check

Scope: Dissolved Oxygen, pH Value, Turbidity, Salinity and Temperature

Brand Name/ Model No.: [YSI]/ [6820-C-M]

Serial No./ Equipment No.: [02J0058-AB]/ [MPP16]

Date of Calibration: 21-November-2023

Ms. Lin Wai Yu, Iris
Assistant Manager - Inorganics

This report may not be reproduced except with prior written approval from ALS Technichem (HK) Pty Ltd.

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION



WORK ORDER: HK2346638
SUB-BATCH: 0
DATE OF ISSUE: 22-Nov-2023
CLIENT: EGS (ASIA) LTD

Equipment Type: Multifunctional Meter
Brand Name/ Model No.: [YSI]/ [6820-C-M]
Serial No./ Equipment No.: [02J0058-AB]/ [MPP16]
Date of Calibration: 21-November-2023

PARAMETERS:

Dissolved Oxygen Method Ref: APHA (23rd edition), 4500O: G

Expected Reading (mg/L)	Displayed Reading (mg/L)	Tolerance (mg/L)
2.94	2.89	-0.05
5.20	5.23	+0.03
7.63	7.68	+0.05
	Tolerance Limit (mg/L)	±0.20

pH Value Method Ref: APHA (23rd edition), 4500H: B

Expected Reading (pH unit)	Displayed Reading (pH unit)	Tolerance (pH unit)
4.0	3.92	-0.08
7.0	6.94	-0.06
10.0	10.15	+0.15
	Tolerance Limit (pH unit)	±0.20

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

Ms. Lin Wai Yu, Iris
Assistant Manager - Inorganics

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION



WORK ORDER: HK2346638
SUB-BATCH: 0
DATE OF ISSUE: 22-Nov-2023
CLIENT: EGS (ASIA) LTD

Equipment Type: Multifunctional Meter
Brand Name/ Model No.: [YSI]/ [6820-C-M]
Serial No./ Equipment No.: [02J0058-AB]/ [MPP16]
Date of Calibration: 21-November-2023

PARAMETERS:

Turbidity

Method Ref: APHA (23rd edition), 2130B

Expected Reading (NTU)	Displayed Reading (NTU)	Tolerance (%)
0	0.0	--
4	3.9	-2.5
40	38.7	-3.2
80	78.6	-1.8
	Tolerance Limit (%)	±10.0

Salinity

Method Ref: APHA (23rd edition), 2520B

Expected Reading (ppt)	Displayed Reading (ppt)	Tolerance (%)
0	0.00	--
10	9.56	-4.4
20	19.04	-4.8
30	28.46	-5.1
	Tolerance Limit (%)	±10.0

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

Ms. Lin Wai Yu, Iris
Assistant Manager - Inorganics

REPORT OF EQUIPMENT PERFORMANCE CHECK/CALIBRATION



WORK ORDER: HK2346638
SUB-BATCH: 0
DATE OF ISSUE: 22-Nov-2023
CLIENT: EGS (ASIA) LTD

Equipment Type: Multifunctional Meter
Brand Name/ Model No.: [YSI]/ [6820-C-M]
Serial No./ Equipment No.: [02J0058-AB]/ [MPP16]
Date of Calibration: 21-November-2023

PARAMETERS:

Temperature

Method Ref: Section 6 of International Accreditation New Zealand Technical Guide No. 3 Second edition March 2008: Working Thermometer Calibration Procedure.

Expected Reading (°C)	Displayed Reading (°C)	Tolerance (°C)
11.5	11.45	-0.1
20.0	19.76	-0.2
40.0	40.02	+0.0
	Tolerance Limit (°C)	±2.0

Remark: "Displayed Reading" presents the figures shown on item under calibration / checking regardless of equipment precision or significant figures.

Ms. Lin Wai Yu, Iris
Assistant Manager - Inorganics